



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**Weighing Technologies, Inc.**  
**2105 Seabrook Circle**  
**Seabrook, TX 77586**  
**(and satellite sites as listed on the scope)**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 31 July 2026

Certificate Number: AC-1112



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Weighing Technologies, Inc.**

2105 Seabrook Circle  
Seabrook, TX 77586

Jodie Stewart  
281-474-5277

**CALIBRATION**

Valid to: **July 31, 2026**

Certificate Number: **AC-1112**

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) <sup>2</sup>	Reference Standard, Method, and/or Equipment
Balances and Scales	1 mg to 500 mg 0.5 g to 2 000 g 2 000 g to 6 000 g	0.012 mg 0.000 45% of reading + 0.05 mg 0.000 26% of reading + 4 mg	Class 1 Weights
	1 mg to 500 mg 0.5 g to 300 g 300 g to 1 000 g 1 000 g to 6 000 g	0.121*W <sup>0.3194</sup> mg 0.024% of reading + 0.86 mg 0.007% of reading + 52 mg 0.012% of reading + 9 mg	Class 5 and Class F Weights
	0.2 lb to 25 000 lb	0.015% of reading + 30 µlb	Class F Weights
	6 000 lb to 200 000 lb	0.014% of reading + 3.2 lb	Class F Weights w/Cart



ANSI National Accreditation Board

### Services performed at satellite laboratory

11475 U.S. HWY 90  
Beaumont, TX 77713  
Jodie Stewart  
281-474-5277

#### Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) <sup>2</sup>	Reference Standard, Method, and/or Equipment
Balances and Scales	1 mg to 500 mg 0.5 g to 2 000 g 2 000 g to 6 000 g	0.012 mg 0.000 45% of reading + 0.05 mg 0.000 25% of reading + 4 mg	Class 1 Weights
	1 mg to 500 mg 0.5 g to 300 g 300 g to 1 000 g 1 000 g to 6 000 g	0.121*W <sup>0.3194</sup> mg 0.024% of reading + 0.86 mg 0.007% of reading + 52 mg 0.012% of reading + 9 mg	Class 5 and Class F Weights
	0.2 lb to 25 000 lb	0.015% of reading + 30 µlb	Class F Weights
	6 000 lb to 200 000 lb	0.014% of reading + 3.2 lb	Class F Weights w/Cart



ANSI National Accreditation Board

### Services performed at satellite laboratory

2422 HWY 288-B  
Richwood, TX 77531  
Jodie Stewart  
281-474-5277

#### Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) <sup>2</sup>	Reference Standard, Method, and/or Equipment
Balances and Scales	1 mg to 500 mg 0.5 g to 2 000 g 2 000 g to 6 000 g	0.012 mg 0.000 45% of reading + 0.05 mg 0.000 25% of reading + 4 mg	Class 1 Weights
	1 mg to 500 mg 0.5 g to 300 g 300 g to 1 000 g 1 000 g to 6 000 g	0.121*W <sup>0.3194</sup> mg 0.024% of reading + 0.86 mg 0.007% of reading + 52 mg 0.012% of reading + 9 mg	Class 5 and Class F Weights
	0.2 lb to 25 000 lb	0.015% of reading + 30 µlb	Class F Weights
	6 000 lb to 200 000 lb	0.014% of reading + 3.2 lb	Class F Weights w/Cart

**Services performed at satellite laboratory**

WTRail  
 2105 Seabrook Circle  
 Seabrook, TX 77586  
 Jodie Stewart  
 281-474-5277

**Mass and Mass Related**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Heavy Capacity Scales <sup>1</sup>	Up to 400 000 lb	0.014% of reading + 3.2 lb	ASTM E617 - Class 7 Test Cart Weights

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. W = Mass weight applied for calibration.
3. Resolution is included in the uncertainty at the time of calibration for the specific device
4. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1112.



Jason Stine, Vice President